

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010835.D
 Acq On : 12 Jul 2019 14:53
 Operator : JC/SP
 Sample : K3782-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC018-SW-190711

Quant Time: Jul 13 03:44:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	200232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	319280	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	296252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	105697	56.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.60%	
35) Dibromofluoromethane	5.50	113	97357	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
50) Toluene-d8	8.71	98	382152	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
62) 4-Bromofluorobenzene	11.13	95	128021	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
Target Compounds						
16) Acetone	2.44	43	5092	5.412	ug/l #	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010835.D
Acq On : 12 Jul 2019 14:53
Operator : JC/SP
Sample : K3782-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC018-SW-190711

Quant Time: Jul 13 03:44:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

