

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010836.D
 Acq On : 12 Jul 2019 15:16
 Operator : JC/SP
 Sample : K3782-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0539-190711

Quant Time: Jul 13 03:47:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	197872	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	312161	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	291127	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132979	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	102851	55.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.88%	
35) Dibromofluoromethane	5.50	113	96507	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
50) Toluene-d8	8.71	98	369756	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	11.13	95	127625	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
Target Compounds						
16) Acetone	2.44	43	3849	4.140	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010836.D
Acq On : 12 Jul 2019 15:16
Operator : JC/SP
Sample : K3782-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0539-190711

Quant Time: Jul 13 03:47:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

