

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010867.D
 Acq On : 13 Jul 2019 03:42
 Operator : JC/SP
 Sample : K3791-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0600

Quant Time: Jul 15 09:10:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	241642	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	381493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	355093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126602	55.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.76%	
35) Dibromofluoromethane	5.50	113	115558	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
50) Toluene-d8	8.71	98	453726	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.13	95	162122	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
Target Compounds						
16) Acetone	2.44	43	3134	2.760	ug/l #	79
31) Cyclohexane	5.59	56	1217	0.372	ug/l #	84
39) Methylcyclohexane	7.46	83	1929	0.484	ug/l	89
84) 1,2,4-Trimethylbenzene	11.80	105	4384	0.464	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010867.D
Acq On : 13 Jul 2019 03:42
Operator : JC/SP
Sample : K3791-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0600

Quant Time: Jul 15 09:10:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

