

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071219\
 Data File : VX010824.D
 Acq On : 12 Jul 2019 10:24
 Operator : JC/SP
 Sample : VX0712WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0712WBL01

Quant Time: Jul 13 01:54:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	206465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	322998	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	290887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126644	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	107743	55.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.32%	
35) Dibromofluoromethane	5.49	113	101802	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
50) Toluene-d8	8.71	98	384471	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
62) 4-Bromofluorobenzene	11.13	95	122140	44.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.50%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071219\
Data File : VX010824.D
Acq On : 12 Jul 2019 10:24
Operator : JC/SP
Sample : VX0712WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0712WBL01

Quant Time: Jul 13 01:54:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010824.D

Acq: 12 Jul 2019 10:24

Instrument :

MSVOA_X

ClientSampleId :


VX0712WBL01

Tot Ion:168 Resp: 206465

Ion Ratio Lower Upper

168 100

99 48.9 35.5 53.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 55.658 ug/l

RT: 6.06 min Scan# 805

Delta R.T. 0.00 min

Lab File: VX010824.D


Acq: 12 Jul 2019 10:24

Tgt Ion: 65 Resp: 107743

Ion Ratio Lower Upper

65 100

67 53.9 0.0 110.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010824.D

Acq: 12 Jul 2019 10:24

Instrument :

MSVOA_X

ClientSampleId :

VX0712WBL01


Tot Ion:114 Resp: 322998

Ion Ratio Lower Upper

114 100

63 18.3 0.0 33.8




88 14.5 0.0 27.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.514 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX010824.D

Acq: 12 Jul 2019 10:24


Tgt Ion:113 Resp: 101802

Ion Ratio Lower Upper

113 100

111 101.0 81.2 121.8



192 22.2 18.6 27.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.852 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010824.D

Acq: 12 Jul 2019 10:24

Instrument :

MSVOA_X

ClientSampleId :





VX0712WBL01

Tot Ion: 98 Resp: 384471

Ion Ratio Lower Upper

98 100

100 63.9 51.6 77.4



#62

4-Bromofluorobenzene

Concen: 44.747 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010824.D

Acq: 12 Jul 2019 10:24




Tgt Ion: 95 Resp: 122140


Ion Ratio Lower Upper

95 100

174 95.7 0.0 187.6

176 92.5 0.0 184.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX010824.D
Acq: 12 Jul 2019 10:24

Instrument :
MSVOA_X
ClientSampled :
VX0712WBL01

Tot Ion:117 Resp: 290887
Ion Ratio Lower Upper
117 100
82 51.1 41.2 61.8
119 32.0 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX010824.D
Acq: 12 Jul 2019 10:24

Tgt Ion:152 Resp: 126644
Ion Ratio Lower Upper
152 100
115 55.0 38.6 115.7
150 157.4 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

