

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071221\  
 Data File : VX023243.D  
 Acq On : 12 Jul 2021 15:48  
 Operator : JC/MD  
 Sample : M2965-01  
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RO-ETGI-345

Quant Time: Jul 12 16:14:34 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 204040   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 328582   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.061 | 117   | 287831   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 122182   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 129956   | 47.390   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 94.780%  |
| 35) Dibromofluoromethane    | 5.397  | 113   | 98981    | 46.649   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 93.300%  |
| 50) Toluene-d8              | 8.653  | 98    | 387087   | 48.906   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 97.820%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 138529   | 45.336   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 90.680%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 44) Trichloroethene         | 7.129  | 130   | 1241     | 0.514    | ug/l   | 93       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071221\  
Data File : VX023243.D  
Acq On : 12 Jul 2021 15:48  
Operator : JC/MD  
Sample : M2965-01  
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA\_X/MEOH  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RO-ETGI-345

Quant Time: Jul 12 16:14:34 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 02 14:29:42 2021  
Response via : Initial Calibration

