

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
 Data File : VX023248.D
 Acq On : 12 Jul 2021 17:45
 Operator : JC/MD
 Sample : M3007-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 13 01:30:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	230757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	377212	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	328331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138337	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155588	50.167	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.340%
35) Dibromofluoromethane	5.397	113	120404	49.429	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	8.653	98	440501	48.479	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.960%
62) 4-Bromofluorobenzene	11.085	95	156698	44.671	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
Data File : VX023248.D
Acq On : 12 Jul 2021 17:45
Operator : JC/MD
Sample : M3007-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jul 13 01:30:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

