

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
 Data File : VX023249.D
 Acq On : 12 Jul 2021 18:09
 Operator : JC/MD
 Sample : M3015-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RED-FRAC-TANK

Quant Time: Jul 13 01:30:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	220040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	360252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	326823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	147351	49.826	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	99.660%	
35) Dibromofluoromethane	5.397	113	113712	48.880	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.760%	
50) Toluene-d8	8.653	98	422726	48.713	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.420%	
62) 4-Bromofluorobenzene	11.085	95	158372	47.273	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.540%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	10227	8.741	ug/l	92
25) 2-Butanone	4.574	43	5681	3.220	ug/l #	88
52) Toluene	8.726	92	1774	0.294	ug/l	97
65) Chlorobenzene	10.085	112	14105	2.183	ug/l	91
83) tert-Butylbenzene	11.719	119	2320	0.282	ug/l	77
84) 1,2,4-Trimethylbenzene	11.756	105	3601	0.420	ug/l	89
85) sec-Butylbenzene	11.896	105	7169	0.677	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
Data File : VX023249.D
Acq On : 12 Jul 2021 18:09
Operator : JC/MD
Sample : M3015-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RED-FRAC-TANK

Quant Time: Jul 13 01:30:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

