

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
 Data File : VX023232.D
 Acq On : 12 Jul 2021 09:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 12 15:43:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.453	0.442	2.4	87	0.00
3 P	Chloromethane	0.448	0.400	10.7	85	0.00
4 C	Vinyl Chloride	0.475	0.436	8.2#	85	0.00
5 T	Bromomethane	0.265	0.221	16.6	85	0.00
6 T	Chloroethane	0.310	0.274	11.6	81	0.00
7 T	Trichlorofluoromethane	0.825	0.734	11.0	81	0.00
8 T	Diethyl Ether	0.305	0.304	0.3	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.481	0.4	94	0.00
10 T	Methyl Iodide	0.593	0.620	-4.6	98	0.00
11 T	Tert butyl alcohol	0.141	0.141	0.0	101	0.00
12 CM	1,1-Dichloroethene	0.446	0.431	3.4#	90	0.00
13 T	Acrolein	0.094	0.088	6.4	90	0.00
14 T	Allyl chloride	0.774	0.787	-1.7	94	0.00
15 T	Acrylonitrile	0.275	0.286	-4.0	97	0.00
16 T	Acetone	0.266	0.265	0.4	99	0.00
17 T	Carbon Disulfide	0.940	0.894	4.9	92	0.00
18 T	Methyl Acetate	0.645	0.657	-1.9	96	0.00
19 T	Methyl tert-butyl Ether	1.667	1.810	-8.6	101	0.00
20 T	Methylene Chloride	0.572	0.531	7.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.481	0.471	2.1	93	0.00
22 T	Diisopropyl ether	1.633	1.614	1.2	92	0.00
23 T	Vinyl Acetate	1.382	1.422	-2.9	93	0.00
24 P	1,1-Dichloroethane	0.923	0.921	0.2	94	0.00
25 T	2-Butanone	0.401	0.408	-1.7	95	0.00
26 T	2,2-Dichloropropane	0.774	0.842	-8.8	101	0.00
27 T	cis-1,2-Dichloroethene	0.583	0.595	-2.1	96	0.00
28 T	Bromochloromethane	0.438	0.414	5.5	91	0.00
29 T	Tetrahydrofuran	0.258	0.253	1.9	91	0.00
30 C	Chloroform	0.997	0.998	-0.1#	94	0.00
31 T	Cyclohexane	0.805	0.720	10.6	86	0.00
32 T	1,1,1-Trichloroethane	0.868	0.879	-1.3	94	0.00
33 S	1,2-Dichloroethane-d4	0.672	0.649	3.4	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.323	0.327	-1.2	102	0.00
36 T	1,1-Dichloropropene	0.446	0.420	5.8	89	0.00
37 T	Ethyl Acetate	0.486	0.484	0.4	93	0.00
38 T	Carbon Tetrachloride	0.465	0.491	-5.6	96	0.00
39 T	Methylcyclohexane	0.521	0.510	2.1	93	0.00
40 TM	Benzene	1.294	1.275	1.5	92	0.00
41 T	Methacrylonitrile	0.272	0.273	-0.4	94	0.00
42 TM	1,2-Dichloroethane	0.520	0.518	0.4	91	0.00
43 T	Isopropyl Acetate	0.809	0.812	-0.4	94	0.00
44 TM	Trichloroethene	0.367	0.375	-2.2	96	0.00
45 C	1,2-Dichloropropane	0.338	0.345	-2.1#	96	0.00
46 T	Dibromomethane	0.241	0.254	-5.4	97	0.00
47 T	Bromodichloromethane	0.441	0.489	-10.9	100	0.00
48 T	Methyl methacrylate	0.397	0.391	1.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
 Data File : VX023232.D
 Acq On : 12 Jul 2021 09:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 12 15:43:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	96	0.00
50 S	Toluene-d8	1.204	1.176	2.3	96	0.00
51 T	4-Methyl-2-Pentanone	0.501	0.499	0.4	91	0.00
52 CM	Toluene	0.838	0.838	0.0	93	0.00
53 T	t-1,3-Dichloropropene	0.481	0.568	-18.1	104	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.588	-15.3	101	0.00
55 T	1,1,2-Trichloroethane	0.352	0.380	-8.0	100	0.00
56 T	Ethyl methacrylate	0.543	0.585	-7.7	99	0.00
57 T	1,3-Dichloropropane	0.581	0.601	-3.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.290	-9.0	93	0.00
59 T	2-Hexanone	0.385	0.382	0.8	90	0.00
60 T	Dibromochloromethane	0.338	0.407	-20.4	107	0.00
61 T	1,2-Dibromoethane	0.361	0.397	-10.0	98	0.00
62 S	4-Bromofluorobenzene	0.465	0.470	-1.1	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.399	0.414	-3.8	96	0.00
65 PM	Chlorobenzene	0.988	1.035	-4.8	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.405	-13.4	100	0.00
67 C	Ethyl Benzene	1.747	1.779	-1.8#	92	0.00
68 T	m/p-Xylenes	0.667	0.699	-4.8	94	0.00
69 T	o-Xylene	0.672	0.703	-4.6	95	0.00
70 T	Styrene	1.098	1.194	-8.7	95	0.00
71 P	Bromoform	0.252	0.324	-28.6#	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.522	3.521	0.0	91	0.00
74 T	N-amyl acetate	1.495	1.465	2.0	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.119	1.129	-0.9	96	0.00
76 T	1,2,3-Trichloropropane	1.006	1.088	-8.2	98	0.00
77 T	Bromobenzene	0.879	0.919	-4.6	97	0.00
78 T	n-propylbenzene	4.073	4.050	0.6	90	0.00
79 T	2-Chlorotoluene	2.517	2.510	0.3	92	0.00
80 T	1,3,5-Trimethylbenzene	3.033	3.098	-2.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.398	-21.3	109	0.00
82 T	4-Chlorotoluene	2.843	2.918	-2.6	93	0.00
83 T	tert-Butylbenzene	2.879	3.005	-4.4	94	0.00
84 T	1,2,4-Trimethylbenzene	2.997	3.107	-3.7	93	0.00
85 T	sec-Butylbenzene	3.704	3.797	-2.5	92	0.00
86 T	p-Isopropyltoluene	3.107	3.254	-4.7	94	0.00
87 T	1,3-Dichlorobenzene	1.651	1.716	-3.9	95	0.00
88 T	1,4-Dichlorobenzene	1.663	1.709	-2.8	95	0.00
89 T	n-Butylbenzene	2.722	2.761	-1.4	90	0.00
90 T	Hexachloroethane	0.455	0.544	-19.6	104	0.00
91 T	1,2-Dichlorobenzene	1.643	1.714	-4.3	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.238	0.278	-16.8	101	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.126	-11.3	102	0.00
94 T	Hexachlorobutadiene	0.428	0.469	-9.6	102	0.00
95 T	Naphthalene	3.366	3.823	-13.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071221\
Data File : VX023232.D
Acq On : 12 Jul 2021 09:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 12 15:43:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.018	1.150	-13.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5