


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071222\
Data File : VX030009.D
Acq On : 12 Jul 2022 09:51
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Title : SW846 8260
Last Update : Wed Jul 13 09:05:23 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.6	9597	PASS
75	95	30	60	46.3	25312	PASS
95	95	100	100	100.0	54664	PASS
96	95	5	9	6.9	3777	PASS
173	174	0.00	2	1.1	451	PASS
174	95	50	100	73.5	40184	PASS
175	174	5	9	7.7	3096	PASS
176	174	95	101	96.6	38803	PASS
177	176	5	9	6.8	2654	PASS