

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
 Data File : VX042299.D
 Acq On : 12 Jul 2024 13:53
 Operator : JC/MD
 Sample : P3220-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 240710064-06-TRIP-BLANK

Quant Time: Jul 15 03:26:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	21866	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	71819	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	50290	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	45343	30.124	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.400%
60) 4-Bromofluorobenzene	11.079	95	18300	23.508	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	78.367%
63) Toluene-d8	8.647	98	66250	30.791	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.633%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
Data File : VX042299.D
Acq On : 12 Jul 2024 13:53
Operator : JC/MD
Sample : P3220-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
240710064-06-TRIP-BLANK

Quant Time: Jul 15 03:26:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

