

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
 Data File : VX042303.D
 Acq On : 12 Jul 2024 15:26
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 15 03:28:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	26314	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	140279	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	124349	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	53645	29.615	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.700%
60) 4-Bromofluorobenzene	11.079	95	51192	26.596	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.667%
63) Toluene-d8	8.647	98	156403	29.399	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
Data File : VX042303.D
Acq On : 12 Jul 2024 15:26
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 15 03:28:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

