

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
 Data File : VX042307.D
 Acq On : 12 Jul 2024 16:58
 Operator : JC/MD
 Sample : PB162021ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB162021ZHE#04

Quant Time: Jul 15 04:45:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158879	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	248963	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84377	46.055	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.120%
35) Dibromofluoromethane	5.385	113	82027	46.434	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	294411	47.440	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.880%
62) 4-Bromofluorobenzene	11.079	95	106629	46.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	24640	30.629	ug/l	# 89
18) Methyl Acetate	2.709	43	2618	1.193	ug/l	98
20) Methylene Chloride	2.788	84	10053	5.707	ug/l	# 85
43) Isopropyl Acetate	6.348	43	108027	28.801	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
Data File : VX042307.D
Acq On : 12 Jul 2024 16:58
Operator : JC/MD
Sample : PB162021ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB162021ZHE#04

Quant Time: Jul 15 04:45:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

