

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
 Data File : VX042309.D
 Acq On : 12 Jul 2024 17:45
 Operator : JC/MD
 Sample : PB162021ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB162021ZHE#06

Quant Time: Jul 15 04:46:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122486	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	205704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79126	56.022	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.040%
35) Dibromofluoromethane	5.385	113	74969	51.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	8.647	98	220600	43.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	86.040%#
62) 4-Bromofluorobenzene	11.079	95	78847	41.715	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.420%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	26217	42.272	ug/l	# 84
18) Methyl Acetate	2.703	43	4060	2.400	ug/l	99
20) Methylene Chloride	2.788	84	7435	5.475	ug/l	90
43) Isopropyl Acetate	6.342	43	108278	34.938	ug/l	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
Data File : VX042309.D
Acq On : 12 Jul 2024 17:45
Operator : JC/MD
Sample : PB162021ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB162021ZHE#06

Quant Time: Jul 15 04:46:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

