

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
 Data File : VX042310.D
 Acq On : 12 Jul 2024 18:08
 Operator : JC/MD
 Sample : PB162021ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB162021ZHE#07

Quant Time: Jul 15 04:47:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146081	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	238055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85802	50.936	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.880%
35) Dibromofluoromethane	5.385	113	82908	49.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.646	98	266139	44.850	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.700%#
62) 4-Bromofluorobenzene	11.079	95	101656	46.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	26182	35.397	ug/l	90
18) Methyl Acetate	2.709	43	3889	1.927	ug/l	92
20) Methylene Chloride	2.788	84	7090	4.378	ug/l	88
43) Isopropyl Acetate	6.342	43	108076	30.134	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071224\
Data File : VX042310.D
Acq On : 12 Jul 2024 18:08
Operator : JC/MD
Sample : PB162021ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB162021ZHE#07

Quant Time: Jul 15 04:47:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

