

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
 Data File : VX003379.D
 Acq On : 13 Jul 2018 23:21
 Operator : JC/MD
 Sample : J3966-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11977

Quant Time: Jul 14 03:26:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 13 17:01:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	399366	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	373260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204913	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176062	59.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.76%	
35) Dibromofluoromethane	5.51	113	152375	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
50) Toluene-d8	8.72	98	570570	59.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.00%	
62) 4-Bromofluorobenzene	11.14	95	199911	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
Target Compounds						
16) Acetone	2.45	43	14302	9.44	ug/l #	82
18) Methyl Acetate	2.78	43	25475	5.42	ug/l	92
20) Methylene Chloride	2.86	84	242781	75.96	ug/l	92
25) 2-Butanone	4.71	43	3199	1.36	ug/l #	90
30) Chloroform	5.21	83	8691	1.49	ug/l	96
95) Naphthalene	13.83	128	201891	14.08	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
Data File : VX003379.D
Acq On : 13 Jul 2018 23:21
Operator : JC/MD
Sample : J3966-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11977

Quant Time: Jul 14 03:26:56 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 13 17:01:02 2018
Response via : Initial Calibration

