

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
 Data File : VX003381.D
 Acq On : 14 Jul 2018 00:14
 Operator : JC/MD
 Sample : J3958-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-CONCRETE

Quant Time: Jul 14 03:36:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 13 17:01:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	296268	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	443234	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204297	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	188832	58.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.90%	
35) Dibromofluoromethane	5.51	113	164507	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	8.72	98	602612	56.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.06%	
62) 4-Bromofluorobenzene	11.14	95	199399	43.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.78%	
Target Compounds						
16) Acetone	2.45	43	15995	9.62	ug/l	93
20) Methylene Chloride	2.86	84	24190	5.66	ug/l	92
25) 2-Butanone	4.71	43	3040	1.18	ug/l	97
43) Isopropyl Acetate	6.47	43	14465	1.79	ug/l	97
74) N-amyl acetate	6.47	43	14465	2.21	ug/l #	98
95) Naphthalene	13.83	128	14956	1.05	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071318\
Data File : VX003381.D
Acq On : 14 Jul 2018 00:14
Operator : JC/MD
Sample : J3958-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-CONCRETE

Quant Time: Jul 14 03:36:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 13 17:01:02 2018
Response via : Initial Calibration

