

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
 Data File : VX003383.D
 Acq On : 14 Jul 2018 01:07
 Operator : JC/MD
 Sample : J3831-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-04-071218

Quant Time: Jul 14 03:38:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 13 17:01:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	280105	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	415174	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	382936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212054	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183007	59.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.96%	
35) Dibromofluoromethane	5.51	113	156749	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
50) Toluene-d8	8.72	98	578728	58.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.00%	
62) 4-Bromofluorobenzene	11.14	95	202038	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
Target Compounds						
16) Acetone	2.45	43	21604	13.74	ug/l	91
18) Methyl Acetate	2.78	43	26237	5.37	ug/l	93
20) Methylene Chloride	2.86	84	127166	37.67	ug/l	94
25) 2-Butanone	4.71	43	3735	1.53	ug/l	98
44) Trichloroethene	7.21	130	9508	2.43	ug/l	94
95) Naphthalene	13.84	128	438042	29.51	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071318\
Data File : VX003383.D
Acq On : 14 Jul 2018 01:07
Operator : JC/MD
Sample : J3831-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-04-071218

Quant Time: Jul 14 03:38:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 13 17:01:02 2018
Response via : Initial Calibration

