

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
 Data File : VX003384.D
 Acq On : 14 Jul 2018 01:33
 Operator : JC/MD
 Sample : PB111071TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB111071TB

Quant Time: Jul 14 03:39:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 13 17:01:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	290243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	419983	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	380535	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209171	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	184478	58.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.56%	
35) Dibromofluoromethane	5.51	113	165335	53.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.98%	
50) Toluene-d8	8.72	98	584206	57.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.76%	
62) 4-Bromofluorobenzene	11.14	95	199350	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
Target Compounds						
16) Acetone	2.45	43	16247	9.97	ug/l	# 88
18) Methyl Acetate	2.78	43	27608	5.46	ug/l	96
20) Methylene Chloride	2.86	84	9709	1.52	ug/l	90
95) Naphthalene	13.84	128	4184	0.29	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
Data File : VX003384.D
Acq On : 14 Jul 2018 01:33
Operator : JC/MD
Sample : PB111071TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB111071TB

Quant Time: Jul 14 03:39:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 13 17:01:02 2018
Response via : Initial Calibration

