

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071318\
 Data File : VX003387.D
 Acq On : 14 Jul 2018 02:53
 Operator : JC/MD
 Sample : J3975-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-6

Quant Time: Jul 14 03:55:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 13 17:01:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	291130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	418975	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	386109	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211251	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	186722	58.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.66%	
35) Dibromofluoromethane	5.51	113	168683	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
50) Toluene-d8	8.72	98	589230	58.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.08%	
62) 4-Bromofluorobenzene	11.14	95	210382	48.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.98%	
Target Compounds						
8) Diethyl Ether	2.19	74	2527	1.05	ug/l	99
16) Acetone	2.45	43	42634	26.09	ug/l	92
18) Methyl Acetate	2.78	43	24860	4.90	ug/l	94
20) Methylene Chloride	2.86	84	121132	34.41	ug/l	86
25) 2-Butanone	4.71	43	12041	4.75	ug/l	95
95) Naphthalene	13.83	128	159925	10.82	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071318\
Data File : VX003387.D
Acq On : 14 Jul 2018 02:53
Operator : JC/MD
Sample : J3975-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-6

Quant Time: Jul 14 03:55:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071318W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 13 17:01:02 2018
Response via : Initial Calibration

