


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071320\
Data File : VX017346.D
Acq On : 13 Jul 2020 09:07
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Title : SW846 8260
Last Update : Tue Jul 14 02:59:51 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	12650	PASS
75	95	30	60	48.5	36598	PASS
95	95	100	100	100.0	75461	PASS
96	95	5	9	6.2	4666	PASS
173	174	0.00	2	0.9	595	PASS
174	95	50	100	87.5	66032	PASS
175	174	5	9	7.6	4989	PASS
176	174	95	101	96.3	63568	PASS
177	176	5	9	6.3	4008	PASS