

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
 Data File : VX023255.D
 Acq On : 13 Jul 2021 11:27
 Operator : JC/MD
 Sample : VX0713MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0713MBL01

Quant Time: Jul 14 04:51:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	227241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	367463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131582	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152771	50.021	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.040%	
35) Dibromofluoromethane	5.397	113	119171	50.221	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.440%	
50) Toluene-d8	8.653	98	434214	49.055	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.120%	
62) 4-Bromofluorobenzene	11.085	95	152994	44.772	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.540%	
Target Compounds						Qvalue
16) Acetone	2.386	43	5494	4.547	ug/l	# 77
30) Chloroform	5.105	83	18790	4.148	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
Data File : VX023255.D
Acq On : 13 Jul 2021 11:27
Operator : JC/MD
Sample : VX0713MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0713MBL01

Quant Time: Jul 14 04:51:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

