

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071321\  
 Data File : VX023261.D  
 Acq On : 13 Jul 2021 14:06  
 Operator : JC/MD  
 Sample : PB137639ZHE#03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB137639ZHE#03

Quant Time: Jul 14 04:52:52 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.562	168	219298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362002	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326175	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137046	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	149447	50.71	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery =	101.42%		
35) Dibromofluoromethane	5.398	113	116162	49.69	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery =	99.38%		
50) Toluene-d8	8.653	98	432613	49.61	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery =	99.22%		
62) 4-Bromofluorobenzene	11.086	95	153197	45.51	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery =	91.02%		
Target Compounds						
16) Acetone	2.386	43	47870	41.05	ug/l	92
20) Methylene Chloride	2.794	84	21210	8.45	ug/l	98
25) 2-Butanone	4.574	43	8792	5.00	ug/l #	88
43) Isopropyl Acetate	6.355	43	30060	5.13	ug/l	94
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071321\  
 Data File : VX023261.D  
 Acq On : 13 Jul 2021 14:06  
 Operator : JC/MD  
 Sample : PB137639ZHE#03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB137639ZHE#03

Quant Time: Jul 14 04:52:52 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X070221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 02 14:29:42 2021  
 Response via : Initial Calibration

