

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
 Data File : VX023262.D
 Acq On : 13 Jul 2021 14:29
 Operator : JC/MD
 Sample : M3012-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PS-TP1

Quant Time: Jul 14 04:53:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.562	168	225083	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	372698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	340031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145912	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	152552	50.43	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	100.86%	
35) Dibromofluoromethane	5.397	113	119991	49.86	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.72%	
50) Toluene-d8	8.653	98	441322	49.16	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.32%	
62) 4-Bromofluorobenzene	11.085	95	162903	47.00	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.00%	
Target Compounds						
16) Acetone	2.386	43	31593	26.40	ug/l	# 87
20) Methylene Chloride	2.794	84	17159	6.66	ug/l	95
43) Isopropyl Acetate	6.348	43	30279	5.02	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
Data File : VX023262.D
Acq On : 13 Jul 2021 14:29
Operator : JC/MD
Sample : M3012-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PS-TP1

Quant Time: Jul 14 04:53:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

