

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
 Data File : VX023264.D
 Acq On : 13 Jul 2021 15:16
 Operator : JC/MD
 Sample : M3013-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-11995

Quant Time: Jul 14 04:53:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	233556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	379904	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	350066	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150178	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156035	49.71	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	99.42%	
35) Dibromofluoromethane	5.397	113	123319	50.27	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.54%	
50) Toluene-d8	8.653	98	459572	50.22	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.085	95	166291	47.07	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.14%	
Target Compounds						
16) Acetone	2.386	43	42595	34.30	ug/l	Qvalue 94
20) Methylene Chloride	2.794	84	17486	6.54	ug/l	98
43) Isopropyl Acetate	6.348	43	30153	4.91	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071321\
Data File : VX023264.D
Acq On : 13 Jul 2021 15:16
Operator : JC/MD
Sample : M3013-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-11995

Quant Time: Jul 14 04:53:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

