

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
 Data File : VX030036.D
 Acq On : 13 Jul 2022 11:47
 Operator : JC/MD
 Sample : N3691-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-02B-18.7-23.7-071222

Quant Time: Jul 14 04:54:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	208119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	362711	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	312766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	114147	45.061	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	90.120%		
35) Dibromofluoromethane	5.385	113	105203	45.402	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.800%		
50) Toluene-d8	8.653	98	413066	47.399	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	94.800%		
62) 4-Bromofluorobenzene	11.079	95	135237	44.718	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.440%		
Target Compounds						
16) Acetone	2.386	43	16475	13.802	ug/l	99
25) 2-Butanone	4.580	43	3659	1.800	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
Data File : VX030036.D
Acq On : 13 Jul 2022 11:47
Operator : JC/MD
Sample : N3691-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-02B-18.7-23.7-071222

Quant Time: Jul 14 04:54:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

