

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
 Data File : VX030040.D
 Acq On : 13 Jul 2022 13:21
 Operator : JC/MD
 Sample : N3695-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-02-07112022

Quant Time: Jul 14 04:55:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	184657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	319275	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	274871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100209	44.585	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.160%
35) Dibromofluoromethane	5.391	113	93883	46.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.060%
50) Toluene-d8	8.653	98	362996	47.320	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.640%
62) 4-Bromofluorobenzene	11.085	95	118335	44.452	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	6743	6.367	ug/l	94
25) 2-Butanone	4.574	43	2870	1.591	ug/l	94
44) Trichloroethene	7.141	130	741	0.306	ug/l	81
52) Toluene	8.726	92	1737	0.330	ug/l	91
64) Tetrachloroethene	9.275	164	8173	3.704	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
Data File : VX030040.D
Acq On : 13 Jul 2022 13:21
Operator : JC/MD
Sample : N3695-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-02-07112022

Quant Time: Jul 14 04:55:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

