

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071322\  
 Data File : VX030041.D  
 Acq On : 13 Jul 2022 13:44  
 Operator : JC/MD  
 Sample : N3695-03  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 EB-01-07052022

Quant Time: Jul 14 04:55:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X071222W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 13 09:05:23 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 179653   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 308400   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 260048   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 102311   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 99528    | 45.515   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 91.040%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 91886    | 46.639   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 93.280%  |
| 50) Toluene-d8              | 8.653  | 98    | 350587   | 47.314   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 94.620%  |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 109069   | 42.416   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 84.840%  |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.386  | 43    | 7011     | 6.804    | ug/l # | 87       |
| 17) Carbon Disulfide        | 2.514  | 76    | 3911     | 0.390    | ug/l   | 97       |
| 25) 2-Butanone              | 4.587  | 43    | 2849     | 1.623    | ug/l # | 76       |
| 30) Chloroform              | 5.117  | 83    | 1044     | 0.302    | ug/l   | 97       |
| 52) Toluene                 | 8.720  | 92    | 1993     | 0.392    | ug/l   | 84       |
| 64) Tetrachloroethene       | 9.275  | 164   | 18944    | 9.075    | ug/l   | 95       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX071322\  
Data File : VX030041.D  
Acq On : 13 Jul 2022 13:44  
Operator : JC/MD  
Sample : N3695-03  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
EB-01-07052022

Quant Time: Jul 14 04:55:38 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X071222W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jul 13 09:05:23 2022  
Response via : Initial Calibration

