

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
 Data File : VX030049.D
 Acq On : 13 Jul 2022 17:09
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 14 04:57:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	167824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	288733	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90819	44.460	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.920%
35) Dibromofluoromethane	5.391	113	84689	45.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.820%
50) Toluene-d8	8.653	98	328425	47.343	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.085	95	111403	46.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	971	0.443	ug/l	84
64) Tetrachloroethene	9.275	164	19276	9.589	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071322\
Data File : VX030049.D
Acq On : 13 Jul 2022 17:09
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 14 04:57:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

