

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX071420\
 Data File : VX017381.D
 Acq On : 14 Jul 2020 16:05
 Operator : JC/SP
 Sample : L3181-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-070920

Quant Time: Jul 15 02:02:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	252501	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	424032	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466761	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	244910	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	160586	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
35) Dibromofluoromethane	5.46	113	141639	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
50) Toluene-d8	8.70	98	538167	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
62) 4-Bromofluorobenzene	11.12	95	229228	56.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.02%	
Target Compounds						
16) Acetone	2.42	43	29078	25.719	ug/l	99
18) Methyl Acetate	2.75	43	3173	1.136	ug/l	99
20) Methylene Chloride	2.84	84	26441	9.585	ug/l	97
43) Isopropyl Acetate	6.42	43	51343	7.966	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX071420\
Data File : VX017381.D
Acq On : 14 Jul 2020 16:05
Operator : JC/SP
Sample : L3181-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
CL-02-070920

Quant Time: Jul 15 02:02:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

