

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071420\
 Data File : VX017365.D
 Acq On : 14 Jul 2020 09:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 01:36:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.423	0.486	-14.9	115	0.00
3 P	Chloromethane	0.476	0.468	1.7	108	0.00
4 C	Vinyl Chloride	0.471	0.475	-0.8#	106	0.00
5 T	Bromomethane	0.307	0.310	-1.0	106	0.00
6 T	Chloroethane	0.297	0.296	0.3	107	0.00
7 T	Trichlorofluoromethane	0.795	0.827	-4.0	108	0.00
8 T	Diethyl Ether	0.281	0.275	2.1	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.451	0.4	106	0.00
10 T	Methyl Iodide	0.550	0.573	-4.2	103	0.00
11 T	Tert butyl alcohol	0.106	0.098	7.5	97	0.00
12 CM	1,1-Dichloroethene	0.455	0.416	8.6#	101	0.00
13 T	Acrolein	0.058	0.050	13.8	88	0.00
14 T	Allyl chloride	0.788	0.785	0.4	108	0.00
15 T	Acrylonitrile	0.256	0.248	3.1	103	0.00
16 T	Acetone	0.224	0.190	15.2	95	0.00
17 T	Carbon Disulfide	1.327	1.326	0.1	108	0.00
18 T	Methyl Acetate	0.553	0.523	5.4	105	0.00
19 T	Methyl tert-butyl Ether	1.545	1.553	-0.5	106	0.00
20 T	Methylene Chloride	0.546	0.507	7.1	107	0.00
21 T	trans-1,2-Dichloroethene	0.521	0.508	2.5	106	0.00
22 T	Diisopropyl ether	1.502	1.523	-1.4	107	0.00
23 T	Vinyl Acetate	1.321	1.353	-2.4	105	0.00
24 P	1,1-Dichloroethane	0.883	0.872	1.2	108	0.00
25 T	2-Butanone	0.356	0.325	8.7	98	0.00
26 T	2,2-Dichloropropane	0.821	0.815	0.7	107	0.00
27 T	cis-1,2-Dichloroethene	0.571	0.552	3.3	105	0.00
28 T	Bromochloromethane	0.408	0.403	1.2	107	0.00
29 T	Tetrahydrofuran	0.230	0.219	4.8	101	0.00
30 C	Chloroform	0.950	0.916	3.6#	105	0.00
31 T	Cyclohexane	0.723	0.761	-5.3	110	0.00
32 T	1,1,1-Trichloroethane	0.878	0.850	3.2	103	0.00
33 S	1,2-Dichloroethane-d4	0.628	0.595	5.3	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	105	0.00
35 S	Dibromofluoromethane	0.337	0.338	-0.3	108	-0.01
36 T	1,1-Dichloropropene	0.460	0.483	-5.0	108	0.00
37 T	Ethyl Acetate	0.476	0.465	2.3	101	0.00
38 T	Carbon Tetrachloride	0.538	0.562	-4.5	107	0.00
39 T	Methylcyclohexane	0.499	0.556	-11.4	111	0.00
40 TM	Benzene	1.322	1.357	-2.6	105	0.00
41 T	Methacrylonitrile	0.255	0.261	-2.4	104	-0.01
42 TM	1,2-Dichloroethane	0.514	0.521	-1.4	105	0.00
43 T	Isopropyl Acetate	0.760	0.756	0.5	101	0.00
44 TM	Trichloroethene	0.404	0.412	-2.0	105	0.00
45 C	1,2-Dichloropropane	0.338	0.353	-4.4#	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071420\
 Data File : VX017365.D
 Acq On : 14 Jul 2020 09:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 01:36:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.244	0.249	-2.0	106	0.00
47 T	Bromodichloromethane	0.510	0.522	-2.4	104	0.00
48 T	Methyl methacrylate	0.365	0.377	-3.3	101	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	95	0.00
50 S	Toluene-d8	1.215	1.245	-2.5	108	0.00
51 T	4-Methyl-2-Pentanone	0.463	0.466	-0.6	100	0.00
52 CM	Toluene	0.848	0.897	-5.8#	106	0.00
53 T	t-1,3-Dichloropropene	0.549	0.568	-3.5	103	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.608	-3.8	104	0.00
55 T	1,1,2-Trichloroethane	0.354	0.364	-2.8	104	0.00
56 T	Ethyl methacrylate	0.502	0.528	-5.2	102	0.00
57 T	1,3-Dichloropropane	0.588	0.599	-1.9	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.255	0.279	-9.4	103	0.00
59 T	2-Hexanone	0.350	0.360	-2.9	101	0.00
60 T	Dibromochloromethane	0.432	0.440	-1.9	103	0.00
61 T	1,2-Dibromoethane	0.385	0.392	-1.8	103	0.00
62 S	4-Bromofluorobenzene	0.483	0.493	-2.1	107	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.433	0.439	-1.4	105	0.00
65 PM	Chlorobenzene	1.016	1.024	-0.8	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.409	-1.0	103	0.00
67 C	Ethyl Benzene	1.692	1.769	-4.6#	104	0.00
68 T	m/p-Xylenes	0.644	0.696	-8.1	105	0.00
69 T	o-Xylene	0.615	0.650	-5.7	103	0.00
70 T	Styrene	1.056	1.152	-9.1	104	0.00
71 P	Bromoform	0.341	0.360	-5.6	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.134	3.366	-7.4	105	0.00
74 T	N-amyl acetate	1.254	1.341	-6.9	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.952	0.946	0.6	102	0.00
76 T	1,2,3-Trichloropropane	0.924	0.951	-2.9	102	0.00
77 T	Bromobenzene	0.893	0.896	-0.3	101	0.00
78 T	n-propylbenzene	3.520	3.866	-9.8	106	0.00
79 T	2-Chlorotoluene	2.169	2.294	-5.8	104	0.00
80 T	1,3,5-Trimethylbenzene	2.661	2.905	-9.2	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.364	0.360	1.1	99	0.00
82 T	4-Chlorotoluene	2.571	2.743	-6.7	105	0.00
83 T	tert-Butylbenzene	2.572	2.742	-6.6	102	0.00
84 T	1,2,4-Trimethylbenzene	2.688	2.902	-8.0	103	0.00
85 T	sec-Butylbenzene	3.011	3.275	-8.8	104	0.00
86 T	p-Isopropyltoluene	2.849	3.167	-11.2	107	0.00
87 T	1,3-Dichlorobenzene	1.597	1.608	-0.7	102	0.00
88 T	1,4-Dichlorobenzene	1.613	1.614	-0.1	103	0.00
89 T	n-Butylbenzene	2.430	2.642	-8.7	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX071420\
Data File : VX017365.D
Acq On : 14 Jul 2020 09:32
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 15 01:36:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.541	0.572	-5.7	105	0.00
91 T	1,2-Dichlorobenzene	1.501	1.501	0.0	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.251	0.230	8.4	99	0.00
93 T	1,2,4-Trichlorobenzene	1.042	1.057	-1.4	103	0.00
94 T	Hexachlorobutadiene	0.425	0.449	-5.6	106	0.00
95 T	Naphthalene	3.063	3.209	-4.8	101	0.00
96 T	1,2,3-Trichlorobenzene	1.006	1.047	-4.1	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6