

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023274.D
 Acq On : 14 Jul 2021 13:13
 Operator : JC/MD
 Sample : M2920-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jul 15 04:26:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	234179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	380459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	334358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139412	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	157288	49.975	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.940%
35) Dibromofluoromethane	5.397	113	122389	49.816	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	8.653	98	447280	48.805	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.085	95	158216	44.718	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.440%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.800	84	2966	1.106	ug/l #	92
95) Naphthalene	13.780	128	1643	1.429	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023274.D
Acq On : 14 Jul 2021 13:13
Operator : JC/MD
Sample : M2920-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jul 15 04:26:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

