

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023276.D
 Acq On : 14 Jul 2021 14:01
 Operator : JC/MD
 Sample : PB137668ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137668ZHE#01

Quant Time: Jul 15 04:26:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	221658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362865	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	321092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	152048	51.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.080%
35) Dibromofluoromethane	5.397	113	116955	49.912	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	8.653	98	430998	49.309	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.085	95	150474	44.592	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.180%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	31081	26.371	ug/l #	85
20) Methylene Chloride	2.794	84	14261	5.620	ug/l #	94
43) Isopropyl Acetate	6.354	43	32037	5.458	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023276.D
Acq On : 14 Jul 2021 14:01
Operator : JC/MD
Sample : PB137668ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137668ZHE#01

Quant Time: Jul 15 04:26:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

