

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023278.D
 Acq On : 14 Jul 2021 14:48
 Operator : JC/MD
 Sample : PB137668ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137668ZHE#03

Quant Time: Jul 15 04:27:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	211546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	344060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	313138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	143656	50.527	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 101.060%		
35) Dibromofluoromethane	5.397	113	111379	50.130	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.260%		
50) Toluene-d8	8.653	98	414211	49.978	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.960%		
62) 4-Bromofluorobenzene	11.085	95	146374	45.748	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.500%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	31251	27.782	ug/l	88
20) Methylene Chloride	2.794	84	13978	5.772	ug/l	90
43) Isopropyl Acetate	6.348	43	31767	5.708	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023278.D
Acq On : 14 Jul 2021 14:48
Operator : JC/MD
Sample : PB137668ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137668ZHE#03

Quant Time: Jul 15 04:27:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

