

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023279.D
 Acq On : 14 Jul 2021 15:11
 Operator : JC/MD
 Sample : PB137668ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137668ZHE#04

Quant Time: Jul 15 05:03:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	218849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	352032	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	320875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	146743	49.890	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.780%
35) Dibromofluoromethane	5.397	113	112492	49.485	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.960%
50) Toluene-d8	8.653	98	424531	50.063	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.120%
62) 4-Bromofluorobenzene	11.085	95	147082	44.928	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.860%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	31484	27.056	ug/l	89
20) Methylene Chloride	2.794	84	15350	6.127	ug/l	88
43) Isopropyl Acetate	6.348	43	31144	5.469	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023279.D
Acq On : 14 Jul 2021 15:11
Operator : JC/MD
Sample : PB137668ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137668ZHE#04

Quant Time: Jul 15 05:03:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

