

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023280.D
 Acq On : 14 Jul 2021 15:35
 Operator : JC/MD
 Sample : PB137668ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137668ZHE#05

Quant Time: Jul 15 04:27:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	227269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	375788	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	331523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	133510	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	156367	51.192	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.380%
35) Dibromofluoromethane	5.397	113	119066	49.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.140%
50) Toluene-d8	8.653	98	445197	49.182	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.360%
62) 4-Bromofluorobenzene	11.085	95	155731	44.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.120%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	34062	28.187	ug/l	91
20) Methylene Chloride	2.794	84	19316	7.424	ug/l	94
43) Isopropyl Acetate	6.360	43	33206	5.463	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023280.D
Acq On : 14 Jul 2021 15:35
Operator : JC/MD
Sample : PB137668ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137668ZHE#05

Quant Time: Jul 15 04:27:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

