

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023281.D
 Acq On : 14 Jul 2021 15:58
 Operator : JC/MD
 Sample : PB137668ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137668ZHE#06

Quant Time: Jul 15 04:28:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	241713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	399866	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	373839	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	166217	51.165	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.340%
35) Dibromofluoromethane	5.391	113	129101	49.997	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.000%
50) Toluene-d8	8.653	98	479922	49.825	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.660%
62) 4-Bromofluorobenzene	11.085	95	178650	48.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	36981	28.773	ug/l	92
20) Methylene Chloride	2.794	84	23427	8.466	ug/l	95
43) Isopropyl Acetate	6.348	43	31163	4.818	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023281.D
Acq On : 14 Jul 2021 15:58
Operator : JC/MD
Sample : PB137668ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137668ZHE#06

Quant Time: Jul 15 04:28:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

