

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023289.D
 Acq On : 14 Jul 2021 19:07
 Operator : JC/MD
 Sample : M3027-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1572

Manual Integrations
 APPROVED

MMDadoda
 7/15/2021 5:51:58 PM

Quant Time: Jul 15 04:30:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	217118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	354593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	324585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	148126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	147378	50.505	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 101.020%			
35) Dibromofluoromethane	5.397	113	111371	48.638	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.280%			
50) Toluene-d8	8.653	98	426527	49.936	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.880%			
62) 4-Bromofluorobenzene	11.085	95	159339	48.321	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.640%			
Target Compounds					Qvalue	
11) Tert butyl alcohol	2.971	59	17269	28.131	ug/l	96
16) Acetone	2.392	43	14313087	12397.930	ug/l #	82
20) Methylene Chloride	2.794	84	23622	9.504	ug/l	90
25) 2-Butanone	4.568	43	37374	21.468	ug/l	93
43) Isopropyl Acetate	6.355	43	47878	8.347	ug/l	93
51) 4-Methyl-2-Pentanone	8.580	43	136205m	38.321	ug/l	
52) Toluene	8.726	92	4217821	710.111	ug/l	94
67) Ethyl Benzene	10.195	91	87806	7.741	ug/l	97
68) m/p-Xylenes	10.305	106	86377	19.946	ug/l	99
69) o-Xylene	10.647	106	61831	14.167	ug/l	99
73) Isopropylbenzene	10.964	105	14095	1.351	ug/l	98
78) n-propylbenzene	11.305	91	38601	3.199	ug/l	100
80) 1,3,5-Trimethylbenzene	11.457	105	55453	6.172	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	253971	28.603	ug/l	99
85) sec-Butylbenzene	11.896	105	11777	1.073	ug/l	92
95) Naphthalene	13.780	128	76509	8.018	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023289.D
Acq On : 14 Jul 2021 19:07
Operator : JC/MD
Sample : M3027-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1572

Manual Integrations
APPROVED

MMDadoda
7/15/2021 5:51:58 PM

Quant Time: Jul 15 04:30:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

