

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
 Data File : VX023290.D
 Acq On : 14 Jul 2021 19:30
 Operator : JC/MD
 Sample : M3021-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31455

Quant Time: Jul 15 04:30:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 02 14:29:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	219379	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	359792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	316495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136409	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	150500	51.044	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.080%		
35) Dibromofluoromethane	5.397	113	115471	49.699	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.400%		
50) Toluene-d8	8.653	98	427752	49.355	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 98.720%		
62) 4-Bromofluorobenzene	11.085	95	152534	45.589	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.180%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	41432	66.796	ug/l	97
16) Acetone	2.386	43	67389	57.770	ug/l	90
20) Methylene Chloride	2.794	84	24195	9.634	ug/l	89
25) 2-Butanone	4.574	43	11531	6.555	ug/l #	83
43) Isopropyl Acetate	6.355	43	31965	5.492	ug/l #	90
52) Toluene	8.726	92	6918	1.148	ug/l	99
95) Naphthalene	13.780	128	2281	1.494	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071421\
Data File : VX023290.D
Acq On : 14 Jul 2021 19:30
Operator : JC/MD
Sample : M3021-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31455

Quant Time: Jul 15 04:30:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 02 14:29:42 2021
Response via : Initial Calibration

