

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030054.D
 Acq On : 14 Jul 2022 12:11
 Operator : JC/MD
 Sample : VX0714WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0714WBL01

Quant Time: Jul 14 12:44:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jul 09 02:10:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	29809	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	163381	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	141021	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	68508	28.185	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	93.933%
60) 4-Bromofluorobenzene	11.085	95	72675	27.346	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.167%
63) Toluene-d8	8.653	98	225373	29.473	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.233%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030054.D
Acq On : 14 Jul 2022 12:11
Operator : JC/MD
Sample : VX0714WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0714WBL01

Quant Time: Jul 14 12:44:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jul 09 02:10:54 2022
Response via : Initial Calibration

