

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030055.D
 Acq On : 14 Jul 2022 12:41
 Operator : JC/MD
 Sample : VX0714WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0714WBL02

Quant Time: Jul 15 05:02:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	180823	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	311652	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109415	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	101229	45.993	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.980%
35) Dibromofluoromethane	5.391	113	93120	46.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.653	98	357857	47.792	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.079	95	112917	43.455	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030055.D
Acq On : 14 Jul 2022 12:41
Operator : JC/MD
Sample : VX0714WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0714WBL02

Quant Time: Jul 15 05:02:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

