

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030058.D
 Acq On : 14 Jul 2022 13:59
 Operator : JC/MD
 Sample : N3672-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-048-IDWGW-20220711

Quant Time: Jul 14 14:40:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jul 09 02:10:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	28368	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	158978	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	138138	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	67728	29.279	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.600%
60) 4-Bromofluorobenzene	11.085	95	69630	26.747	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	89.167%
63) Toluene-d8	8.653	98	221353	29.551	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.500%
Target Compounds						
					Qvalue	
24) Methyl tert-Butyl Ether	3.117	73	8633	1.340	ug/l	94
61) Tetrachloroethene	9.281	164	2816	1.358	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030058.D
Acq On : 14 Jul 2022 13:59
Operator : JC/MD
Sample : N3672-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-048-IDWGW-20220711

Quant Time: Jul 14 14:40:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jul 09 02:10:54 2022
Response via : Initial Calibration

