

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030059.D
 Acq On : 14 Jul 2022 14:22
 Operator : JC/MD
 Sample : N3672-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20220711

Quant Time: Jul 14 14:49:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jul 09 02:10:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	27912	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	154207	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	133788	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	67482	29.649	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.833%
60) 4-Bromofluorobenzene	11.085	95	70227	27.853	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.833%
63) Toluene-d8	8.653	98	213261	29.397	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.000%
Target Compounds						
					Qvalue	
23) tert-Butyl Alcohol	2.947	59	9749	15.446	ug/l	# 100
61) Tetrachloroethene	9.275	164	2347	1.168	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030059.D
Acq On : 14 Jul 2022 14:22
Operator : JC/MD
Sample : N3672-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20220711

Quant Time: Jul 14 14:49:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X070822W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jul 09 02:10:54 2022
Response via : Initial Calibration

