

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030064.D
 Acq On : 14 Jul 2022 16:33
 Operator : JC/MD
 Sample : N3705-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-149S1-20220712

Quant Time: Jul 15 04:18:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	168873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	291216	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94800	46.120	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.240%		
35) Dibromofluoromethane	5.391	113	85966	46.208	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.420%		
50) Toluene-d8	8.653	98	327603	46.821	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.640%		
62) 4-Bromofluorobenzene	11.085	95	105235	43.340	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.680%		
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	3.117	73	70783	13.015	ug/l	98
22) Diisopropyl ether	3.769	45	4977	0.846	ug/l #	77
24) 1,1-Dichloroethane	3.611	63	2759	0.846	ug/l #	83
27) cis-1,2-Dichloroethene	4.495	96	1436	0.654	ug/l	95
42) 1,2-Dichloroethane	6.098	62	2871	1.225	ug/l	95
44) Trichloroethene	7.128	130	18024	8.155	ug/l	94
64) Tetrachloroethene	9.274	164	5397	2.705	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030064.D
Acq On : 14 Jul 2022 16:33
Operator : JC/MD
Sample : N3705-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-149S1-20220712

Quant Time: Jul 15 04:18:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

