

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030067.D
 Acq On : 14 Jul 2022 17:44
 Operator : JC/MD
 Sample : N3695-04DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01C-84DL

Quant Time: Jul 15 04:19:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	286347	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92283	46.437	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.880%
35) Dibromofluoromethane	5.397	113	84290	46.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.653	98	322810	46.921	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.840%
62) 4-Bromofluorobenzene	11.085	95	102827	43.069	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.140%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	2349	2.508	ug/l #	85
27) cis-1,2-Dichloroethene	4.501	96	33574	15.812	ug/l	88
44) Trichloroethene	7.135	130	4560	2.098	ug/l	93
64) Tetrachloroethene	9.274	164	70291	36.229	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030067.D
Acq On : 14 Jul 2022 17:44
Operator : JC/MD
Sample : N3695-04DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01C-84DL

Quant Time: Jul 15 04:19:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

