

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030069.D
 Acq On : 14 Jul 2022 18:31
 Operator : JC/MD
 Sample : N3695-06DL 500X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01E-84DL

Quant Time: Jul 15 04:19:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	168759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	296871	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96417	46.939	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.880%
35) Dibromofluoromethane	5.397	113	88250	46.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.653	98	336894	47.232	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.460%
62) 4-Bromofluorobenzene	11.085	95	111951	45.228	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.460%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	2678	1.189	ug/l	85
64) Tetrachloroethene	9.274	164	89252	42.705	ug/l	98
90) Hexachloroethane	12.542	117	1657	1.493	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030069.D
Acq On : 14 Jul 2022 18:31
Operator : JC/MD
Sample : N3695-06DL 500X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01E-84DL

Quant Time: Jul 15 04:19:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

