

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030070.D
 Acq On : 14 Jul 2022 18:54
 Operator : JC/MD
 Sample : N3718-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220713004-03-TRIP-BLANK

Quant Time: Jul 15 04:19:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	296693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96159	47.073	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.140%
35) Dibromofluoromethane	5.397	113	87460	46.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.280%
50) Toluene-d8	8.653	98	342489	48.045	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.100%
62) 4-Bromofluorobenzene	11.085	95	107857	43.600	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.200%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	13893	14.433	ug/l	93
30) Chloroform	5.098	83	2470	0.765	ug/l #	71
44) Trichloroethene	7.122	130	593	0.263	ug/l	21
64) Tetrachloroethene	9.274	164	2659	1.294	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030070.D
Acq On : 14 Jul 2022 18:54
Operator : JC/MD
Sample : N3718-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220713004-03-TRIP-BLANK

Quant Time: Jul 15 04:19:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

