

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030078.D
 Acq On : 14 Jul 2022 22:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 04:21:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.530	0.637	-20.2	87	0.00
3 P	Chloromethane	0.521	0.596	-14.4	86	0.00
4 C	Vinyl Chloride	0.579	0.657	-13.5#	88	0.00
5 T	Bromomethane	0.246	0.284	-15.4	103	0.00
6 T	Chloroethane	0.326	0.358	-9.8	87	0.00
7 T	Trichlorofluoromethane	0.747	0.869	-16.3	90	0.00
8 T	Diethyl Ether	0.309	0.342	-10.7	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.510	0.561	-10.0	89	0.00
10 T	Methyl Iodide	0.419	0.623	-48.7#	100	0.00
11 T	Tert butyl alcohol	0.157	0.154	1.9	86	0.00
12 CM	1,1-Dichloroethene	0.509	0.585	-14.9#	90	0.00
13 T	Acrolein	0.031	0.026	16.1	74	0.00
14 T	Allyl chloride	0.890	0.992	-11.5	88	0.00
15 T	Acrylonitrile	0.357	0.413	-15.7	92	0.00
16 T	Acetone	0.287	0.332	-15.7	96	0.00
17 T	Carbon Disulfide	1.282	1.459	-13.8	80	0.00
18 T	Methyl Acetate	0.872	0.985	-13.0	96	0.00
19 T	Methyl tert-butyl Ether	1.610	1.892	-17.5	93	0.00
20 T	Methylene Chloride	0.663	0.673	-1.5	90	0.00
21 T	trans-1,2-Dichloroethene	0.531	0.625	-17.7	91	0.00
22 T	Diisopropyl ether	1.743	2.082	-19.4	95	0.00
23 T	Vinyl Acetate	1.475	1.818	-23.3	93	0.00
24 P	1,1-Dichloroethane	0.966	1.117	-15.6	93	0.00
25 T	2-Butanone	0.488	0.573	-17.4	94	0.00
26 T	2,2-Dichloropropane	0.678	0.566	16.5	67	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.726	-11.7	89	0.00
28 T	Bromochloromethane	0.406	0.459	-13.1	91	0.00
29 T	Tetrahydrofuran	0.332	0.387	-16.6	94	0.00
30 C	Chloroform	0.962	1.129	-17.4#	94	0.00
31 T	Cyclohexane	0.882	1.015	-15.1	90	0.00
32 T	1,1,1-Trichloroethane	0.786	0.920	-17.0	92	0.00
33 S	1,2-Dichloroethane-d4	0.609	0.618	-1.5	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.319	0.307	3.8	84	0.00
36 T	1,1-Dichloropropene	0.413	0.457	-10.7	92	0.00
37 T	Ethyl Acetate	0.530	0.598	-12.8	94	0.00
38 T	Carbon Tetrachloride	0.385	0.416	-8.1	89	0.00
39 T	Methylcyclohexane	0.523	0.580	-10.9	88	0.00
40 TM	Benzene	1.299	1.455	-12.0	91	0.00
41 T	Methacrylonitrile	0.285	0.329	-15.4	93	0.00
42 TM	1,2-Dichloroethane	0.402	0.468	-16.4	97	0.00
43 T	Isopropyl Acetate	0.797	0.927	-16.3	96	0.00
44 TM	Trichloroethene	0.379	0.400	-5.5	90	0.00
45 C	1,2-Dichloropropane	0.338	0.380	-12.4#	94	0.00
46 T	Dibromomethane	0.232	0.260	-12.1	92	0.00
47 T	Bromodichloromethane	0.422	0.471	-11.6	93	0.00
48 T	Methyl methacrylate	0.368	0.435	-18.2	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
 Data File : VX030078.D
 Acq On : 14 Jul 2022 22:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 04:21:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	90	0.00
50 S	Toluene-d8	1.201	1.184	1.4	85	0.00
51 T	4-Methyl-2-Pentanone	0.528	0.611	-15.7	95	0.00
52 CM	Toluene	0.825	0.913	-10.7#	92	0.00
53 T	t-1,3-Dichloropropene	0.438	0.481	-9.8	87	0.00
54 T	cis-1,3-Dichloropropene	0.502	0.533	-6.2	86	0.00
55 T	1,1,2-Trichloroethane	0.340	0.381	-12.1	93	0.00
56 T	Ethyl methacrylate	0.511	0.597	-16.8	92	0.00
57 T	1,3-Dichloropropane	0.540	0.615	-13.9	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.305	-9.7	86	0.00
59 T	2-Hexanone	0.405	0.479	-18.3	95	0.00
60 T	Dibromochloromethane	0.331	0.363	-9.7	90	0.00
61 T	1,2-Dibromoethane	0.354	0.394	-11.3	91	0.00
62 S	4-Bromofluorobenzene	0.417	0.423	-1.4	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.401	0.438	-9.2	91	0.00
65 PM	Chlorobenzene	0.996	1.070	-7.4	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.339	0.380	-12.1	93	0.00
67 C	Ethyl Benzene	1.687	1.935	-14.7#	92	0.00
68 T	m/p-Xylenes	0.662	0.744	-12.4	91	0.00
69 T	o-Xylene	0.663	0.739	-11.5	91	0.00
70 T	Styrene	1.100	1.238	-12.5	92	0.00
71 P	Bromoform	0.273	0.282	-3.3	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.481	3.884	-11.6	92	0.00
74 T	N-amyl acetate	1.529	1.765	-15.4	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.256	1.327	-5.7	93	0.00
76 T	1,2,3-Trichloropropane	1.150	1.253	-9.0	94	0.00
77 T	Bromobenzene	0.871	0.902	-3.6	89	0.00
78 T	n-propylbenzene	4.130	4.624	-12.0	92	0.00
79 T	2-Chlorotoluene	2.474	2.712	-9.6	94	0.00
80 T	1,3,5-Trimethylbenzene	2.934	3.236	-10.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.398	0.386	3.0	80	0.00
82 T	4-Chlorotoluene	2.772	3.052	-10.1	93	0.00
83 T	tert-Butylbenzene	2.813	3.052	-8.5	92	0.00
84 T	1,2,4-Trimethylbenzene	2.902	3.276	-12.9	93	0.00
85 T	sec-Butylbenzene	3.719	4.185	-12.5	93	0.00
86 T	p-Isopropyltoluene	2.984	3.360	-12.6	92	0.00
87 T	1,3-Dichlorobenzene	1.609	1.726	-7.3	90	0.00
88 T	1,4-Dichlorobenzene	1.652	1.763	-6.7	92	0.00
89 T	n-Butylbenzene	2.673	3.211	-20.1	95	0.00
90 T	Hexachloroethane	0.510	0.563	-10.4	90	0.00
91 T	1,2-Dichlorobenzene	1.635	1.738	-6.3	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.330	-15.0	95	0.00
93 T	1,2,4-Trichlorobenzene	1.004	1.040	-3.6	87	0.00
94 T	Hexachlorobutadiene	0.401	0.389	3.0	84	0.00
95 T	Naphthalene	3.987	4.348	-9.1	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071422\
Data File : VX030078.D
Acq On : 14 Jul 2022 22:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 15 04:21:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.022	1.075	-5.2	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6