

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046973.D
 Acq On : 14 Jul 2025 15:02
 Operator : JC/MD
 Sample : Q2565-04
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0196

Quant Time: Jul 15 01:21:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/15/2025
 Supervised By :Semsettin Yesilyurt 07/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	311535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	535193	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	489066	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	243403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	216796	52.676	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.360%			
35) Dibromofluoromethane	5.397	113	168652	46.423	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.840%			
50) Toluene-d8	8.647	98	624812	48.746	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.500%			
62) 4-Bromofluorobenzene	11.079	95	251366	51.600	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.200%			
Target Compounds						
16) Acetone	2.410	43	7789	6.093	ug/l	93
18) Methyl Acetate	2.721	43	11939	3.543	ug/l #	84
20) Methylene Chloride	2.794	84	5465	1.446	ug/l #	82
52) Toluene	8.720	92	298598	32.502	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	50409	3.587	ug/l	96
89) n-Butylbenzene	12.329	91	19793m	1.389	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046973.D
Acq On : 14 Jul 2025 15:02
Operator : JC/MD
Sample : Q2565-04
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0196

Quant Time: Jul 15 01:21:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/15/2025
Supervised By :Semsettin Yesilyurt 07/15/2025

