

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046974.D
 Acq On : 14 Jul 2025 15:23
 Operator : JC/MD
 Sample : Q2565-05
 Misc : 2.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MOO-25-0180

Quant Time: Jul 15 01:21:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	299313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	529387	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	458887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	213116	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	218861	55.349	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.700%
35) Dibromofluoromethane	5.397	113	170792	47.528	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.060%
50) Toluene-d8	8.647	98	591256	46.634	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.260%
62) 4-Bromofluorobenzene	11.079	95	226077	46.918	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.840%
Target Compounds						
						Qvalue
16) Acetone	2.422	43	6357	5.176	ug/l	# 83
20) Methylene Chloride	2.794	84	5946	1.637	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046974.D
Acq On : 14 Jul 2025 15:23
Operator : JC/MD
Sample : Q2565-05
Misc : 2.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MOO-25-0180

Quant Time: Jul 15 01:21:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

