

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
 Data File : VX046979.D
 Acq On : 14 Jul 2025 17:10
 Operator : JC/MD
 Sample : Q2578-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-A2-11-G

Quant Time: Jul 15 01:24:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	327424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	571871	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	520250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	268456	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	235554	54.456	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.920%
35) Dibromofluoromethane	5.562	113	557	0.143	ug/l	0.16
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.280%#
50) Toluene-d8	8.653	98	657591	48.013	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.079	95	270900	52.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.080%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	109109	81.205	ug/l	99
20) Methylene Chloride	2.806	84	21366	5.378	ug/l	98
95) Naphthalene	13.774	128	1785469	117.074	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX071425\
Data File : VX046979.D
Acq On : 14 Jul 2025 17:10
Operator : JC/MD
Sample : Q2578-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-A2-11-G

Quant Time: Jul 15 01:24:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

